

HP 33311A/B/C, 33312B, 33313A/B/C Coaxial Switches

Product Overview


dc to 4 GHz
dc to 18 GHz
dc to 26.5 GHz

High performance switches for microwave and RF instrumentation and systems

Hewlett-Packard offers a versatile line of multiport coaxial switches. These 50Ω mechanical-latching switches come with the performance and reliability that you have come to expect in HP microwave accessories.

Offering excellent electrical and mechanical performance with broadband operation, high isolation, low SWR, long life and exceptional repeatability, these switches are designed with your applications in mind. Frequency ranges are available to 4 GHz and 18 GHz for the A and B models with SMA connectors and to 26.5 GHz for the C models with -3.5 mm connectors.

The HP 33311A, HP 33311B, and HP 33311C (Figure 1) are single-pole double-throw switches with high isolation, >100 dB at 4 GHz for the A model, >90 dB at 18 GHz for the B model and >50 dB at 26.5 GHz for the C model. All models terminate the unused port with 50Ω, a key feature for your applications where low SWR is required on all ports.



The HP 33312B (Figure 2) has 4 RF ports with one internal 50Ω termination designed for applications requiring a transfer switch or a cross switching element.




Figure 1. HP 33311




Figure 2. HP 33312B

COPYRIGHT AND DISCLAIMER NOTICE

Copyright – Agilent Technologies, Inc. Reproduced with the permission of Agilent Technologies Inc. Agilent Technologies, Inc. makes no warranty of any kind with regard to this material including, but not limited to, the implied warranties of merchantability and fitness for a particular purpose. Agilent Technologies, Inc. is not liable for errors contained herein or for incidental or consequential damages in connection with the furnishing, performance, or use of this material or data.

The HP 33313A, 33313B, and 33313C (Figure 3) round out the family with 5 RF ports, giving you flexibility to configure the switch for your specific needs.




Figure 3. HP 33313A/B/C

Applications

Multi-source switching

The HP 33311A, 33311B or 33311C is an excellent choice for applications where you require selection of multiple signal sources, frequency counters, or signal control devices such as modulators or filters, or routing signals from multiple sources as shown in Figure 4.




Figure 4. Multi-source switching

Transfer switching

For applications requiring a "transfer switch," as shown in Figure 5, an HP 33312B is the ideal choice to insert a component, such as a filter or mixer, into a signal path. Another popular use is to switch between a device under test and a through path for system calibration. The HP 33312B's internal load can terminate the device under test when in the through mode (up to 1 watt).




Figure 5. HP 33312B used as a transfer switch (equivalent circuit action)




Figure 6. Signal reversal

Reverse signal path

Figure 6 shows how an HP 33311 and HP 33313 may be used to provide a "transceiver" configuration. This example illustrates how one amplifier can be used to transmit or receive. Any application where the signal path direction needs to be reversed could use this configuration.

Portable and remote applications

Due to their small package size, light weight, low power consumption, and high reliability, these switches are ideal for your portable or remote applications. The light weight and repeatability make them well suited for portable spectrum analyzers and other portable microwave test equipment. Their 1,000,000 cycle life make them the ideal candidate for system design, such as communications repeaters or remote monitoring stations that demand minimal maintenance. The switch's automatic coil disconnect feature and low current drain during switching minimize the amount of power needed for operating the switches.

Dedicated Switching

For larger switching systems, where many switches will be used to provide complex signal routing, a switch driver such as the HP 87130A or 70611A is recommended. The HP 87130A rack-and-stack switch driver and the MMS-based HP 70611A Option 001 are convenient, flexible HP-Interface Bus (HP-IB) or Modular Systems Interface Bus (MS-IB) switch controllers, providing driver circuitry, indicator readback circuitry and firmware that makes it easy to integrate switch components into a switching system. Controlling the HP 87130A is simple using either a PC or workstation based HP-IB controller and HP Interactive Test Generator (HP-ITG) or HP-Visual Engineering Environment (HP-VEE). The HP 70611A gives manual control via the MMS user interface or can also be controlled via an HP-IB equipped PC or workstation.

In addition, the built-in firmware makes it possible to define often used switch paths. With the path command, macros can be designed which open and close the right solenoids to select the desired switch port, and the path may be given a meaningful name. Remember that only one select pin should be activated at one time to prevent rapid cycling of the switch.

Both the HP 87130A and 70611A provide position monitoring and reporting, which make it possible for a program to determine if all the switches are in their proper state (position) before the program continues with testing.

A programmable wake up condition makes it possible to ensure that the matrix or switching system starts up in a predetermined state, to prevent damage to delicate equipment from excessive power. This would also be the state that the system returns to after a power interruption.

Reference literature number 5091-3268E, *HP 87130A Attenuator/Switch Driver* and literature number 5952-3715E, *HP 70611A Attenuator/Switch Driver*.

Accessory cables and adapters make it easy to quickly get the *HP 33311/12/13 working with the HP 87130A or HP 70611A*. Reference literature number 5963-2038E, *Switch Attenuator Driver Configuration Guide*.

For smaller switching needs, the HP 11713A attenuator/switch controller provides simple HP-IB control for up to ten HP 33311/12/13 switches. Connecting cables can be ordered which make it easy to connect the switches to the HP 11713A.

Help for your special needs

For more information on the applications of Hewlett-Packard Switches, request Application Note 332, *Microwave Switching from SPDT to Full-Access Matrix*, literature number 5953-6466, or contact HP for your special switching needs.

Application Note 332-1, *Novel Combinations of Microwave Switches and Step Attenuators for Programming Applications*, literature number 5954-8892, expands on some interesting product capabilities.

Operation and use

How the switch works

All the switches are “break before make”; the switched ports are not connected to each other preventing possible damage to sensitive circuits. The standard configuration switch uses 24 Vdc for the switching voltage, Option 011 specifies a voltage of 5 Vdc, and Option 015 specifies 15 Vdc.

Driving the switch

For standard, Option 011, and 015 switches, switching is accomplished by applying the supply voltage to pin “C” and grounding the appropriate pin to actuate the switching mechanism. **Avoid grounding both RF path select pins as rapid cycling may occur.** After the switch is fully latched (30 ms), the drive current is automatically disconnected. If the drive circuit is pulsed, the pulse duration must be at least 30 ms to ensure that the switch will fully latch.

Remote Indication

The position of the switch may be determined by utilizing the open and closed states of the internal coil contacts. Figure 7 displays two indicator circuits, one to provide a TTL output and one that directly activates an LED. The circuits will output a TTL “HI” and LED “ON” if the switch is in the state shown in Figure 7. When drive pin 1 contact is closed (as shown), RF port 2 is connected to common port. For the HP 33311/12/13, see Figures 2, 3, and Table 1 for corresponding RF port connections. The circuits shown are designed to operate with coil voltage of 24V. If other voltage switches are used, the circuit components must be modified to accommodate that voltage.

Since current is drawn through the solenoid for these indicator circuits, inadvertent switching is prevented by limiting the current to 5 mA. Additional design information may be required from the component manufacturer for the circuits described in Figure 7, depending upon specific applications. *Hewlett-Packard assumes no responsibility for use of any circuits described herein, and makes no representations or warranties, expressed or implied, that such circuits are free from patent infringements.*




Figure 7. TTL and LED remote indicator circuits

Model Number	<i>Standard Drive</i>		RF Path
	Pin 1	Pin 2	
HP 33311A/B/C	Ground	Open	1 to C closed 2 terminated
	Open	Ground	2 to C closed 1 terminated
HP 33312B	Ground	Open	1 to 2 closed 3 to 4 closed
	Open	Ground	1 terminated 2 to 3 closed 4 open
HP 33313A/B/C	Ground	Open	1 open 2 to 3 closed 4 to 5 closed
	Open	Ground	1 to 2 closed 3 to 4 closed 5 open

Table 1. Switching logic table

Specifications

HP model numbers	HP 33311A HP 33313A	HP 33311B HP 33312B HP 33313B	HP 33311C HP 33313C
Frequency range	dc to 4 GHz	dc to 18 GHz	dc to 26.5 GHz
Insertion loss	<0.20 dB, dc to 2 GHz <0.25 dB, 2 to 4 GHz	<0.20 dB, dc to 2 GHz <0.50 dB, 2 to 18 GHz	<0.25 dB, dc to 2 GHz <0.50 dB, 2 to 18 GHz <1.25 dB, 18 to 26.5 GHz
Isolation between ports >50 dB, 18 to 26.5 GHz	>100 dB, dc to 4 GHz	>90 dB, dc to 18 GHz	>90 dB, dc to 18 GHz
SWR through line	<1.10, dc to 2 GHz <1.20, 2 to 4 GHz	<1.10, dc to 2 GHz <1.20, 2 to 12.4 GHz <1.30, 12.4 to 18 GHz	<1.15, dc to 2 GHz <1.25, 2 to 12.4 GHz <1.40, 12.4 to 18 GHz <1.80, 18 to 26.5 GHz
Into internal 50Ω load (HP 8762s and HP 8763s)	<1.10, dc to 2 GHz <1.20, 2 to 4	<1.10, dc to 2 GHz <1.20, 2 to 12.4 GHz <1.30, 12.4 to 18 GHz	<1.15, dc to 2 GHz <1.25, 2 to 12.4 GHz <1.30, 12.4 to 18 GHz <1.80, 18 to 26.5 GHz
Connectors	SMA (f)	SMA (f)	3.5 mm (f)
Insertion loss repeatability dc to 18 GHz 18 to 26.5 GHz (Up to 1,000,000 cycles measured at 25° C)	0.03 dB maximum	0.03 dB maximum	0.03 dB maximum 0.5 dB maximum

General operating data

Maximum power rating	1 watt average; 100 watts peak (not to exceed average power rating); +7 Vdc
Life	1,000,000 cycles minimum
Switching solenoid	Switching speed 30 ms maximum

Switch drive data

Parameter	Test conditions	Minimum	Nominal	Maximum	Units
Supply voltage					
Standard		20	24	32	V
Option 015		12	15	20	V
Option 011		4.5	5	7	V
Supply current	Switching; pulse width ≥ 30 ms maximum				
Standard	@ Vcc = 24 Vdc		120		mA
Option 015	@ Vcc = 15 Vdc		182		mA
Option 011	@ Vcc = 5 Vdc		400		mA
Impedance					
Standard			200		ohm
			127		mH
Option 015			82		ohm
			57		mH
Option 011			13		ohm
			8		mH

Environmental

Operating temperature -25° to +75°C	Humidity 95% RH, 40°C, 5 days	Vibration 0.05", 10 to 55 Hz	Shock 50 g's, 3 ms
--	-------------------------------------	------------------------------------	--------------------------

Physical specifications

Dimensions:	Per Figure 9
Weight:	245 g (9 oz.)

Specifications (continued)




Figure 9. Product outlines

For more information about Hewlett-Packard test and measurement products, applications, services, and for a current sales office listing, visit our web site, <http://www.hp.com/go/tmdir>. You can also contact one of the following centers and ask for a test and measurement sales representative.

United States:

Hewlett-Packard Company
Test and Measurement Call Center
P.O. Box 4026
Englewood, CO 80155-4026
1 800 452 4844

Canada:

Hewlett-Packard Canada Ltd.
5150 Spectrum Way
Mississauga, Ontario L4W 5G1
(905) 206 4725

Europe:

Hewlett-Packard
European Marketing Centre
P.O. Box 999
1180 AZ Amstelveen
The Netherlands
(31 20) 547 9900

Japan:

Hewlett-Packard Japan Ltd.
Measurement Assistance Center
9-1, Takakura-Cho, Hachioji-Shi,
Tokyo 192, Japan
Tel: (81) 426-56-7832
Fax: (81) 426-56-7840

Latin America:

Hewlett-Packard
Latin American Region Headquarters
5200 Blue Lagoon Drive, 9th Floor
Miami, Florida 33126, U.S.A.
(305) 267 4245/4220

Australia/New Zealand:

Hewlett-Packard Australia Ltd.
31-41 Joseph Street
Blackburn, Victoria 3130, Australia
1 800 629 485

Asia Pacific:

Hewlett-Packard Asia Pacific Ltd.
17-21/F Shell Tower, Times Square,
1 Matheson Street, Causeway Bay,
Hong Kong
Tel: (852) 2599 7777
Fax: (852) 2506 9285

Technical Data Sheet Supplement



**HEWLETT
PACKARD**

**HP Part No. 5951-5993
Printed in USA January 1995**

Edition 2

© Copyright 1994 Hewlett-Packard Company. All Rights Reserved. Reproduction, adaptation, or translation without prior written permission is prohibited, except as allowed under the copyright laws.

CERTIFICATION

Hewlett-Packard Company certifies that this product met its published specifications at the time of shipment from the factory. Hewlett-Packard further certifies that its calibration measurements are traceable to the United States National Institute of Standards and Technology (NIST, formerly NBS), to the extent allowed by the institute's calibration facility, and to the calibration facilities of other International Standards Organization members.

WARRANTY

This Hewlett-Packard instrument product is warranted against defects in material and workmanship for a period of one year from date of delivery. During the warranty period, Hewlett-Packard Company will, at its option, either repair or replace products which prove to be defective.

For warranty service or repair, this product must be returned to a service facility designated by HP. Buyer shall prepay shipping charges to HP and HP shall pay shipping charges to return the product to Buyer. However, Buyer shall pay all shipping charges, duties, and taxes for products returned to HP from another country.

HP warrants that its software and firmware designated by HP for use with an instrument will execute its programming instructions when properly installed on that instrument. HP does not warrant that the operation of the instrument, or software, or firmware will be uninterrupted or error free.

LIMITATION OF WARRANTY

The foregoing warranty shall not apply to defects resulting from improper maintenance by Buyer, Buyer-supplied software or interfacing, unauthorized modification or misuse, operation outside of the environmental specifications for the product, or improper site preparation or maintenance.

NO OTHER WARRANTY IS EXPRESSED OR IMPLIED. HP SPECIFICALLY DISCLAIMS THE IMPLIED WARRANTIES OR MERCHANTABILITY AND FITNESS FOR A PARTICULAR PURPOSE.

EXCLUSIVE REMEDIES

THE REMEDIES PROVIDED HEREIN ARE BUYER'S SOLE AND EXCLUSIVE REMEDIES. HP SHALL NOT BE LIABLE FOR ANY DIRECT, INDIRECT, SPECIAL, INCIDENTAL, OR CONSEQUENTIAL DAMAGES, WHETHER BASED ON CONTRACT, TORT, OR ANY OTHER LEGAL THEORY.

ASSISTANCE

Product maintenance agreements and other customer assistance agreements are available for Hewlett-Packard Products.

For any assistance, contact your nearest Hewlett-Packard Sales and Service Office. Addresses are provided at the back of this manual.

General Information

This information sheet supplements the instruments Technical Data Sheet. The data sheet describes the instrument, its specifications, options, and accessories. The instrument is certified and warranted as indicated in this document. For further information contact the nearest Hewlett-Packard Sales and Service offices (see Sales and Service section).

Initial Inspection

Inspect the shipping container for damage. If the shipping container or packing material is damaged it should be kept until the contents of the shipment have been checked mechanically and electrically. If there is mechanical damage or if the instrument does not meet its specifications, notify the nearest Hewlett-Packard office. Keep the damaged shipping materials (if any) for the carrier and a Hewlett-Packard representative to inspect.

Packaging

The following general instructions should be used for re-packaging with commercially available materials:

- a. Wrap the instrument in heavy paper or plastic. If shipping to a Hewlett-Packard office or service center, attach a tag indicating the type of service required, return address, model number, and full serial number.
- b. Use a strong shipping container. A single-wall corrugated carton made of 200 pound-per-square-inch test material is adequate.
- c. Use enough shock-absorbing material around all sides of the instrument to provide firm cushion and prevent movement inside the container.
- d. Seal the shipping container securely.
- e. Mark the shipping container FRAGILE to assure careful handling.

Environment

Non-Operating

The instrument should be stored in a clean, dry environment. The following environmental limitations apply to both storage and shipment (unless otherwise stated in the data sheet):

- Temperature, -55 to $+75^{\circ}\text{C}$
- Humidity, less than 95% relative at $+40^{\circ}\text{C}$
- Altitude, less than 15,300 metres (50,000 feet)

Operating

The operating environment of the instrument should be within the following limitations (unless otherwise stated in the data sheet):

- Temperature, 0 to +55°C
- Humidity, less than 95% relative at +40°C
- Altitude, less than 4,600 metres (15,000 feet)

Note Storage or operation of the instrument within an environment other than that which is specified may cause damage to the instrument and may void the warranty.

Adjustments

No adjustments should be attempted.

Repair

If the instrument fails to operate within the specifications listed in the data sheet, it should be returned to Hewlett-Packard for repair.

Note Attempting to service or repair the instrument may void the warranty.

Hewlett-Packard Sales and Service Offices

IN THE UNITED STATES

California

Hewlett-Packard Co.
1421 South Manhattan Ave.
P.O. Box 4230
Fullerton, CA 92631
(714) 999-6700

Hewlett-Packard Co.
301 E. Evelyn
Mountain View, CA 94039
(415) 694-2000

Colorado

Hewlett-Packard Co.
24 Inverness Place, East
Englewood, CO 80112
(303) 649-5000

Georgia

Hewlett-Packard Co.
2000 South Park Place
P.O. Box 105005
Atlanta, GA 30339
(404) 955-1500

Illinois

Hewlett-Packard Co.
5201 Tollview Drive
Rolling Meadows, IL 60008
(312) 255-9800

New Jersey

Hewlett-Packard Co.
120 W. Century Road
Paramus, NJ 07653
(201) 265-5000

Texas

Hewlett-Packard Co.
930 E. Campbell Rd.
Richardson, TX 75081
(214) 231-6101

IN AUSTRALIA

Hewlett-Packard Australia Ltd.
31-41 Joseph Street
Blackburn, Victoria 3130
895-2895

IN CANADA

Hewlett-Packard (Canada) Ltd.
17500 South Service Road
Trans-Canada Highway
Kirkland, Quebec H9J 2X8
(514) 697-4232

IN FRANCE

Hewlett-Packard France
F-91947 Les Ulis Cedex
Orsay
(6) 907-78-25

**IN GERMAN FEDERAL
REPUBLIC**

Hewlett-Packard GmbH
Vertriebszentrale Frankfurt
Berner Strasse 117
Postfach 560 140
D-6000 Frankfurt 56
(0611) 50-04-1

IN GREAT BRITAIN

Hewlett-Packard Ltd.
King Street Lane
Winnersh, Wokingham
Berkshire RG11 5AR
0734 784774

**IN OTHER EUROPEAN
COUNTRIES**

Hewlett-Packard (Schweiz) AG
Allmend 2
CH-8967 Widen (Zurich)
(0041) 57 31 21 11

IN JAPAN

Yokogawa-Hewlett-Packard Ltd.
29-21 Takaido-Higashi, 3 Chome
Suginami-ku Tokyo 168
(03) 331-6111

**IN PEOPLE'S REPUBLIC
OF CHINA**

China Hewlett-Packard, Ltd.
P.O. Box 9610, Beijing
4th Floor, 2nd Watch Factory
Main Bldg.
Shuang Yu Shu, Bei San Huan Rd.
Beijing, PRC
256-6888

IN SINGAPORE

Hewlett-Packard Singapore
Pte. Ltd.
1150 Depot Road
Singapore 0410
273 7388
Telex HPSGSO RS34209
Fax (65) 2788990

IN TAIWAN

Hewlett-Packard Taiwan
8th Floor, Hewlett-Packard
Building
337 Fu Hsing North Road
Taipei
(02) 712-0404

IN ALL OTHER LOCATIONS

Hewlett-Packard Inter-Americas
3495 Deer Creek Rd.
Palo Alto, California 94304

HP

TECHNICAL DATA SHEET SUPPLEMENT

PRINTED: JAN 1995

EDITION:

HP PART NUMBER : 5951-5993

